

## **Solution Log Sheet**

Location/ Dept	
AER Model/ Serial #	

Strip Lot #:	Date Test Strip Bottle First Opened	Do Not Use After (6 months after opened)
QC Test Date	QC Test Results	Tested by (initials)

Date Solution Opened	Date Solution Expires (75 days from opening)	Date Solution Tested	Time Solution Tested	Temp in °F Before Use	MRC Test Results (circle one)	Tested By (initials)	Comments
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		
					Pass Fail		

NOTE: Do not use Rapicide OPA/28 High-Level Disinfectant beyond its stated use & reuse life (28 days) even if MRC level is met. Use test strips within 180 days of opening bottle.



10 Ranick Rd. Hauppauge, NY 11788-4209 Tel: 888-276-7783 Fax: 631-582-1726

crosstex.com



6789 West Hennetta Rd. Rush, NY 14543 Tel: 800-722-1529 Fax: 585-359-1726 spsmedical.com