In-Office Biological Monitoring System - Record Keeper

Systéme de surveillance biologique - Journal d'enregistrement

Initials *Initiales*	Sterilizer Number & Type *Numéro et type de sterilizateur*	Date & Time into Incubator *Date et heure dans l'incubateur*	Date & Time out of Incubator *Date et heure á l'exterieur de l'incubateur*	* Results (Circle One) ** Résultats (En encercler un)*	** Control (Circle One) *** Contrôle (En encercler un)*
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)
				PASS (Purple) FAIL (Yellow)	PASS (Yellow) FAIL (Purple)

* A sterilized biological indicator should remain purple after incubation, indicating adequate sterilization.
** The control biological indicator should turn yellow, indicating spore growth and a viable test.

** Un bioindicateur stérilisé doit demeurer violet aprés l'incubation, ce qui indique qu'il y a eu une stérilisation adéquate.*
*** Le bioindicateur de contrôle doit devenir jaune, ce qui indique qu'il y a un développement des spores et un test viable.*

Customer Care
1-800-819-3336
customercare@crosstex.com
*Visit **cdc.gov** for additional information*

Infection Control Helpline
TOLL FREE (1-855-878-3745)
1-8558-STERILE

CROSSTEX A CANTEL MEDICAL COMPANY | PROTECTS®