

How often should I perform biological indicator monitoring (spore testing)?

The CDC recommends testing **at least weekly** with a biological monitor for each sterilizer to ensure proper functioning. Users should follow the manufacturer's directions concerning the appropriate placement of the BI in the sterilizer. In addition to conducting routine biological monitoring, equipment users should perform biological monitoring:

- · Whenever a new type of packaging material or tray is used.
- · After training new sterilization personnel.
- · After any sterilizer repair or maintenance.
- · After any change in the sterilizer loading procedures.

Circle the date each week that BI testing was performed

JANUARY

3	IVI		VV		г.	3
	1	2	3	4	5	6
7	8	9	10	11	12	13
14	15	16	17	18	19	20
21	22	23	24	25	26	27
28	29	30	31			

FEBRUARY

S	M	T	W	T	F	S
				1	2	3
4	5	6	7	8	9	10
11	12	13	14	15	16	17
18	19	20	21	22	23	24
25	26	27	28	29		

MARCH

S	M	T	W	T	F	S
					1	2
3	4	5	6	7	8	9
10	11	12	13	14	15	16
17	18	19	20	21	22	23
24	25	26	27	28	29	30
31						

TEST DATE	INITIALS OF PERSON WHO RAN TEST	MAIL DATE	INITIALS OF PERSON WHO MAILED TEST

FOR STERILITY ASSURANCE:

EVERY LOAD: Sure-Check[™] Sterilization Pouches **EVERY DAY:** SteamPlus[™] Integrating Indicators

EVERY WEEK: ConFirm Biological Monitoring Services

For test status or questions call customer care at: 1-800-722-1529 | CrosstexBMS.com

 $\hbox{Hu-Friedy Mfg. Co., LLC, 1666 E. Touhy Ave., Des Plaines, IL 60018}$

 $All\ company\ and\ product\ names\ are\ trademarks\ of\ Hu-Friedy\ Mfg.\ Co.,\ LLC\ its\ affiliates\ or\ related\ companies,\ unless\ otherwise\ noted$





How often should I perform biological indicator monitoring (spore testing)?

The CDC recommends testing **at least weekly** with a biological monitor for each sterilizer to ensure proper functioning. Users should follow the manufacturer's directions concerning the appropriate placement of the BI in the sterilizer. In addition to conducting routine biological monitoring, equipment users should perform biological monitoring:

- · Whenever a new type of packaging material or tray is used.
- · After training new sterilization personnel.
- · After any sterilizer repair or maintenance.
- · After any change in the sterilizer loading procedures.

Circle the date each week that BI testing was performed

APRIL

3	IVI		VV			3
	1	2	3	4	5	6
7	8	9	10	11	12	13
14	15	16	17	18	19	20
21	22	23	24	25	26	27
28	29	30				

MAY

S	M	T	W	T	F	S
			1	2	3	4
5	6	7	8	9	10	11
12	13	14	15	16	17	18
19	20	21	22	23	24	25
26	27	28	29	30	31	

JUNE

S	M	T	W	T	F	S
						1
2	3	4	5	6	7	8
9	10	11	12	13	14	15
16	17	18	19	20	21	22
23	24	25	26	27	28	29
30						

TEST DATE	INITIALS OF PERSON WHO RANTEST	MAIL DATE	INITIALS OF PERSON WHO MAILED TEST

FOR STERILITY ASSURANCE:

EVERY LOAD: Sure-Check Sterilization Pouches **EVERY DAY:** SteamPlus Integrating Indicators

EVERY WEEK: ConFirm Biological Monitoring Services

For test status or questions call customer care at: 1-800-722-1529 | CrosstexBMS.com

 $\hbox{Hu-Friedy Mfg. Co., LLC, 1666 E. Touhy Ave., Des Plaines, IL 60018}$

 $All\ company\ and\ product\ names\ are\ trademarks\ of\ Hu-Friedy\ Mfg.\ Co.,\ LLC\ its\ affiliates\ or\ related\ companies,\ unless\ otherwise\ noted$







How often should I perform biological indicator monitoring (spore testing)?

The CDC recommends testing **at least weekly** with a biological monitor for each sterilizer to ensure proper functioning. Users should follow the manufacturer's directions concerning the appropriate placement of the BI in the sterilizer. In addition to conducting routine biological monitoring, equipment users should perform biological monitoring:

- · Whenever a new type of packaging material or tray is used.
- · After training new sterilization personnel.
- · After any sterilizer repair or maintenance.
- · After any change in the sterilizer loading procedures.

Circle the date each week that BI testing was performed

JULY

S	M	T	W	T	F	S
	1	2	3	4	5	6
7	8	9	10	11	12	13
14	15	16	17	18	19	20
21	22	23	24	25	26	27
28	29	30	31			

AUGUST

S	M	Т	W	Т	F	S
				1	2	3
4	5	6	7	8	9	10
11	12	13	14	15	16	17
18	19	20	21	22	23	24
25	26	27	28	29	30	31

SEPTEMBER

S	M	T	W	T	F	S
1	2	3	4	5	6	7
8	9	10	11	12	13	14
15	16	17	18	19	20	21
22	23	24	25	26	27	28
29	30					

TEST DATE	INITIALS OF PERSON WHO RAN TEST	MAIL DATE	INITIALS OF PERSON WHO MAILED TEST

FOR STERILITY ASSURANCE:

EVERY LOAD: Sure-Check Sterilization Pouches **EVERY DAY:** SteamPlus Integrating Indicators

EVERY WEEK: ConFirm Biological Monitoring Services

For test status or questions call customer care at: 1-800-722-1529 | CrosstexBMS.com

 $\hbox{Hu-Friedy Mfg. Co., LLC, 1666 E. Touhy Ave., Des Plaines, IL 60018}$

 $All\ company\ and\ product\ names\ are\ trademarks\ of\ Hu-Friedy\ Mfg.\ Co.,\ LLC\ its\ affiliates\ or\ related\ companies,\ unless\ otherwise\ noted$





How often should I perform biological indicator monitoring (spore testing)?

The CDC recommends testing **at least weekly** with a biological monitor for each sterilizer to ensure proper functioning. Users should follow the manufacturer's directions concerning the appropriate placement of the BI in the sterilizer. In addition to conducting routine biological monitoring, equipment users should perform biological monitoring:

- · Whenever a new type of packaging material or tray is used.
- · After training new sterilization personnel.
- · After any sterilizer repair or maintenance.
- · After any change in the sterilizer loading procedures.

Circle the date each week that BI testing was performed

OCTOBER

3	IVI		VV			3
		1	2	3	4	5
6	7	8	9	10	11	12
13	14	15	16	17	18	19
20	21	22	23	24	25	26
27	28	29	30	31		

NOVEMBER

S	M	T	W	T	F	S
					1	2
3	4	5	6	7	8	9
10	11	12	13	14	15	16
17	18	19	20	21	22	23
24	25	26	27	28	29	30

DECEMBER

S	M	T	W	T	F	S
1	2	3	4	5	6	7
8	9	10	11	12	13	14
15	16	17	18	19	20	21
22	23	24	25	26	27	28
29	30	31				

TEST DATE	INITIALS OF PERSON WHO RAN TEST	MAIL DATE	INITIALS OF PERSON WHO MAILED TEST

FOR STERILITY ASSURANCE:

EVERY LOAD: Sure-Check Sterilization Pouches **EVERY DAY:** SteamPlus Integrating Indicators

EVERY WEEK: ConFirm Biological Monitoring Services

For test status or questions call customer care at: 1-800-722-1529 | CrosstexBMS.com

 $\hbox{Hu-Friedy Mfg. Co., LLC, 1666 E. Touhy Ave., Des Plaines, IL 60018}$

 $All\ company\ and\ product\ names\ are\ trademarks\ of\ Hu-Friedy\ Mfg.\ Co.,\ LLC\ its\ affiliates\ or\ related\ companies,\ unless\ otherwise\ noted$

